

HYGISON®

REF HS0501A



FFP2 Mask

Non Medical

EN149: 2001 + A1: 2009

FFP2-MASKE

Nicht medizinisch EN149:2001+A1:2009

HYCISUN[®]

REF | HS0501A

Faltbare Partikel -Atemschutzmaske

Hohe Filtrationseffizienz
Geringer Atemwiderstand
Bequem zu tragen



20 Stück

Ⓢ Einmalgebrauch


CE 2797

20 Stück

Ⓢ Einmalgebrauch

CE 2797

Bequem zu tragen
Geringer Atemwiderstand



Bedienungsanleitung Fitting instructions



1 Nehmen Sie die Maske heraus und öffnen Sie beide Seiten.
Take out the mask and open both sides.



2 Drücken Sie die Atemschutzmaske mit dem Bügel auf dem Nasenrücken gegen Ihr Gesicht. Legen Sie das obere Band hoch auf Ihren Hinterkopf. Bewegen Sie das untere Band über Ihren Kopf und positionieren Sie es hinter Ihren Ohren.
Press the mask against your face with the nose clip on the bridge of your nose. Place the top tie high on the back of your head. Move the lower tie over your head and position it behind your ears.



3 Formen Sie den Bügel mit beiden Händen in die Form Ihrer Nase.
Shape the nose clip into the shape of your nose with both hands.



4 Testen Sie die Passform. Nehmen Sie beide Hände über die Atemschutzmaske und atmen Sie kräftig aus. Wenn Luft um Ihre Nase strömt, ziehen Sie den Bügel fester.
Test the correct fit. Put both hands over the respirator and exhale forcefully. When air flows out around your nose, press the nose clip tighter.

Hinweis zur Verwendung:

- Bitte verwenden Sie dieses Produkt nicht in der Nähe einer Feuerquelle.
- Aufgrund persönlicher Unterschiede kann Nebel in den Augen auftreten. Achten Sie daher beim Fahren besser darauf.
- Da es sich bei diesem Produkt um eine Einwegmaske handelt, kann es nicht durch Waschen wiederverwendet werden.
- Von hohen Temperaturen und Luftfeuchtigkeit fernhalten und an einem sauberen Ort aufbewahren.
- Verwenden Sie einzeln verpackte Produkte, sobald diese ausgepackt sind.

NOTICE FOR USE:

- Please do not use this product near fire sources.
- Due to personal differences, fog may appear in the eyes, so please pay more attention when driving.
- As this product is a disposable mask, it cannot be reused through washing.
- Keep it away from high temperature and humidity, and keep it in a clean place.
- Use individually packaged products as soon as they are unpacked.

Einsatzbereich / RANGE OF APPLICATIONS:

Geeignet zum Schutz gegen Pulver, PM2.5-Dunst, Bakterien, Speichel, Staub usw. Die Gesichtsmaske ist winddicht und schützt das Gesicht vor Kälte.

suitable for protection against powder, PM2.5 smog, particulates, flu, bacteria, saliva, dust etc. The face mask is good for wind proof and protection the face against cold weather.

Vorsichtsmaßnahmen :

1. Halten Sie die Gesichtsmaske vor dem Gebrauch sauber und berühren Sie nicht die Innenseite.
2. Stellen Sie sicher, dass die Gesichtsmaske zum Gesicht passt, indem Sie das Nasenstück einstellen und die Maskenkante nahe am Gesicht halten.
3. Überprüfen Sie die Luftdichtung der Maske, verwenden Sie diese nicht, wenn die Gesichtsmaske gebrochen ist, ein Loch oder eine Undichtigkeit aufweist.
4. Waschen Sie die Gesichtsmaske nicht mit Wasser, da Wasser die Filterfähigkeit und den Schutz beeinträchtigt.
5. Dunsten Sie die Gesichtsmaske nicht und stellen Sie sie nicht in die Mikrowelle.
6. Halten Sie die Maske von Dämpfen, Ölen, Chemikalien, Farben, Flüssigkeiten, Säuren und Laugen fern.
7. Entsorgen Sie die Gesichtsmaske, wenn sie ziemlich schmutzig wird oder der Atemwiderstand offensichtlich zunimmt.


Tragemethode / WEARING METHOD:

Öffnen Sie die Maske und legen Sie es auf den Nasenrücken an. Wenn Sie die Maske angezogen haben, drücken Sie auf den Nasenrücken, um sie zu schließen.

Open the mask and place the bridge of the nose on it. After wearing the mask, press the bridge of the nose to close it. The protective effect is better.

CAUTIONS:

1. Keep the face mask clean before use, do not touch the inner side of facemask.
2. Ensure the face mask fit to face, by adjusting the nosepiece, keeping mask edge close to face.
3. Check the air sealing of mask, do not use if the facemask is broken or with hole or leakage.
4. Do not wash the facemask with water, water will destroy filtering ability and protection.
5. Do not steam the facemask, and do not put it in microwave.
6. Keep the face mask away from vapor, oils, chemicals, paints, liquid, acid, alkaline objects.
7. Discard the face mask when it gets quite dirty or breathing resistance increase obviously.





FFP2 MASK - 20 PCS
Mask Mate - 6 PCS
Instruction - 1 PCS
Qualified Certificate - 1 PCS



EU-KONFORMITÄTSERKLÄRUNG

Diese Konformitätserklärung wurde unter der alleinigen Verantwortung des Herstellers

Hunan Dreaming Cloud E-Commerce CO., Ltd.

Block 1, Smart Tech Park, 57 # Huangxing Avenue, Changsha Economic and Technological Development Zone, Changsha, Hunan, China

ausgestellt.

EG-Vertreter: Sunbeam International GmbH, Schumanstr.12, Würselen 52146 Deutschland

Hiermit wird erklärt, dass die folgende persönliche Schutzausrüstung (PSA)

Produktbeschreibung: HYGISUN Partikelfilter-Halbmaske

Produktmodell (e): HS0501A FFP2 NR ohne Ventil

den Bestimmungen der folgenden europäischen Verordnung entspricht:

PSA-Verordnung (Persönliche Schutzausrüstung)

Das Modell entspricht den Bestimmungen der Verordnung (EU) 2016/425, PSA zur Verwendung durch Angehörige der Gesundheitsberufe gemäß der Empfehlung der Kommission 2020/403 und der Nationalen Norm zur Umsetzung der harmonisierten europäischen Normnummer (n):

EN 149: 2001 + A1: 2009

und ist identisch mit der PSA, die Gegenstand einer EU-Typprüfung ist (Modul B der Verordnung (EU) 2016/425), auf die auf der Zertifikatsnummer verwiesen wird:

Zertifikat Nr.: CE 730303 (Ausstellungsdatum: 03/07/2020)

herausgegeben von BSI Group Niederlande BV

John M. Keynesplein 9, 1066 EP, Amsterdam, Niederlande (Notified Body No. 2797)

und entspricht den Verfahren in Modul C2 der Verordnung (EU) 2016/425 unter der Überwachung der BSI Group The Netherlands BV (Notified Body Nr. 2797), auf die auf dem vom BSI ausgestellten Zertifikat CE 730304 (Ausstellungsdatum: 03/07/2020) verwiesen wird.

Changsha, China, 07.04.2020

Ouyang Zhouya
(Nachname Name)

Qualitätsmanager

Hunan Dreaming Cloud E-Commerce CO., Ltd.

EU DECLARATION OF CONFORMITY

This Declaration of Conformity, issued under the sole responsibility of the manufacturer

Hunan Dreaming Cloud E-Commerce CO., Ltd.

Block 1, Smart Tech Park, 57# Huangxing Avenue, Changsha Economic and Technological
Development Zone, Changsha, Hunan, China

EC Representative: Sunbeam International GmbH, Schumanstr.12, Würselen 52146 Germany:

hereby declaring the following Personal Protective Equipment (PPE)

Product Description: HYGISUN Particulate Filtering Half Mask

Product Model/s: HS0501A FFP2 NR without valve

is/are in conformity with the provisions of the following European Regulation

PPE (Personal Protective Equipment) Regulation

The model is/are in conformity with the provisions of Regulation (EU) 2016/425, PPE for use by
healthcare professionals as per Commission recommendation 2020/403, and with the National
Standard transposing the harmonised European Standard Number(s):

EN 149:2001+A1:2009

and is/are identical to the PPE which is/are the subject of EU type-examination (Module B of
Regulation (EU) 2016/425) referenced on the certificate number:

Certificate No.: CE 730303 (Issue Date: 03/07/2020)

issued by

BSI Group The Netherlands B.V.

John M. Keynesplein 9, 1066 EP, Amsterdam, Netherlands (Notified Body No. 2797)

and is/are in conformance with procedures set out in Module C2 of Regulation (EU) 2016/425 under
the surveillance of BSI Group The Netherlands B.V.(Notified Body No. 2797), referenced on BSI
issued Certificate CE 730304 (Issue Date: 03/07/2020).

Changsha, China, 04/07/2020

Ou Yang Zhou

(Surname, Name)

Quality Manager

Hunan Dreaming Cloud E-Commerce CO., Ltd.



EU Type Examination Certificate

This is to certify that:

Sunbeam International GmbH
Schumanstr. 12
Würselen
52146
Germany

Holds Certificate Number:

CE 730303

In respect of:

Model HYGISUN HS0501A Face mask
To technical specification Annex II (EHSR) of the PPE Regulation (EU) 2016/425
PPE for use by healthcare professionals as per Commission recommendation 2020/403


on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Previous Notified Body: BSI 0086

First Issued: 2020-07-03

Latest Issue: 2020-07-03



Drs. Dave Hagenaaars, Managing Director

Effective Date: 2020-07-03

Expiry Date: 2021-07-03

Page: 1 of 3

EU Type Examination Certificate

No. CE 730303

Product Specification

Product Name: Particulate Respirator.

Product Type: Particulate filtering half masks for use by Healthcare professionals.

Model: **HYGISUN HS0501A.**

Classification: FFP2 NR un-valved.

Technical Specification: Technical specification to satisfy Annex II of the PPE Regulation (EU) 2016/425.

Product Description: The respirator is non-reusable, secured to the face of the user by a pair of elasticated ear straps, and has no exhalation valve. The respirator is FFP2 class, vertical fold flat type.

The respirator listed on this certificate is for use by healthcare workers, first responders and other personnel involved in the efforts to contain the COVID-19 virus and avoid its further spread.

The product covered by this certificate is not approved for industrial applications and the certificate is only valid as long as EU Commission recommendation sheet 2020/403 remains applicable.

Product Assessments: BSI's PPE for Healthcare Professionals 2020/403 – RPE Technical Specification.

First Issued: 2020-07-03

Latest Issue: 2020-07-03

Effective Date: 2020-07-03

Expiry Date: 2021-07-03

Page: 2 of 3

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 730303

Certificate Administration Details

Technical File Reference: Sunbeam International GmbH, TCF.01, V0 dated 28/06/2020.

Certificate Amendment Record:

Issue date	Comments	BSI Review No.
July 2020	First issue.	2797:20:3220783

Certificate validity

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall process utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The validity of the Certificate for the products is also dependent on the maintenance of the EU Conformity to Type based on Internal Production Control plus supervised product checks at random intervals, Annex VII (Module C2) as referenced on BSI issued Certificate CE 730304.

First Issued: 2020-07-03

Latest Issue: 2020-07-03

Effective Date: 2020-07-03

Expiry Date: 2021-07-03

Page: 3 of 3

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

Conformity to Type based on Internal Production Control plus supervised product checks at random intervals

This is to certify that:

Sunbeam International GmbH
Schumanstr. 12
Würselen
52146
Germany

Holds Certificate Number:

CE 730304

In respect of:

For the manufacture of particulate respirators to technical specification to satisfy Annex II of the PPE Regulation (EU) 2016/425.

on the basis that BSI carried out the supervised production checks at random intervals under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex VII (Module C2)

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Previous Notified Body: BSI 0086

First Issued: 2020-07-03

Latest Issue: 2020-07-03

Drs. Dave Hagenaaars, Managing Director

Effective Date: 2020-07-03

Expiry Date: 2021-07-03

Page: 1 of 3



...making excellence a habit.™

Conformity to Type based on Internal Production Control plus supervised product checks at random intervals

No. CE 730304

Product manufactured by:

Hunan Dreaming Cloud E-Commerce CO., Ltd
Block 1, Smart Tech Park,
57# Huangxing Avenue,
Changsha Economic and Technological Development Zone,
Changsha,
Hunan,
China

Product details

The respiratory protective device covered by the scope of this Module C2 Certificate and the Technical Specification to which the product is manufactured are as follows:

Product type:	Particulate filtering half masks for use by Healthcare professionals.
Model and classifications:	HYGISUN HS0501A FFP2 NR
Technical Specification:	Technical specification to satisfy Annex II of the PPE Regulation (EU) 2016/425. BSI's PPE for Healthcare Professionals 2020/403 – RPE Technical Specification.

First Issued: 2020-07-03

Latest Issue: 2020-07-03

Effective Date: 2020-07-03

Expiry Date: 2021-07-03

Page: 2 of 3

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

Conformity to Type based on Internal Production Control plus supervised product checks at random intervals

No. CE 730304

Certificate Administration Details:

Certificate Amendment Record and BSI internal Review relating to this Certificate

Issue date	Comments	BSI Review No.
July 2020	First issue.	2797:20:3220784

Certificate validity

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspects of the overall quality system utilized in the manufacture of the products, failure to do so could invalidate the Certificate in respect of product manufactured after the introduction of such changes.

First Issued: 2020-07-03

Latest Issue: 2020-07-03

Effective Date: 2020-07-03

Expiry Date: 2021-07-03

Page: 3 of 3

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.